

Summary of Laser Service Subcontractor Policies at DOE Facilities
 Laser EFCOG Working Group Report by M. Woods (SLAC LSO), October 11, 2009

	Policy Item	Yes	No	Not stated	Situation dependent	Lab manager determines	Subcontract determines	SLAC Policy
1	Is service subcontractor policy referenced in lab's laser safety policy?	5	5					yes
2	Is evidence of training required for service subcontractors?	4	5		1			yes
3	Are specific safety documents describing the work required?	5	3			2		yes
4	Is a Notice-to-proceed required from purchasing department prior to service work?	1	7	2				yes
5	Is LSO approval required?	5	5					yes
5	Is a pre-job briefing required?	4	3			3		yes
6	Lab general safety orientation training required?	2	5*		3			no*
7	Can work alone in laser lab?	3	1		3	3		no
8	Is alignment eyewear allowed?	6	0	1	1	1	1	yes
9	Can alignment eyewear be used for pulsed Ti:sapphire lasers at 800nm?	1	2	5		1	1	yes
10	If alignment eyewear is allowed, is special approval needed?	5	1	3		1		yes
11	Is a CoHE or LOTO evaluation done?	1	7**				1	yes

*not if escorted (for some labs)

**may be done, but not as part of laser safety policy (for some labs); for some labs, only consider this for non-laser hazards